



Tricefy Uplink Requirements

Last Modified on 2024-01-24 13:51

Tricefy Uplink has minimal requirements:

1. An account with [Tricefy](#) (user name and password)
2. Internet connection with "broadband speed" or better.
 - **Tip:** Faster internet connections will result in faster uploads
3. DICOM - licensed ultrasound system(s) connected to an active network
 - The ultrasound system and Uplink software host must be able to talk to each other over the network
4. A computer, server, or virtual machine for running the software. Tricefy Uplink can be run in conjunction with other applications and does not need a dedicated computer.
 - Windows Requirements:
 - Windows 10 or later | Windows Server 2016 or later (**previous versions are no longer supported**)
 - 100 MB of disk space necessary for uplink assets and temporary file processing
 - [Static IP Address](#)
 - Mac OS Requirements:
 - macOS 10.14 (Mojave) or later
 - 100 MB of disk space necessary for uplink assets and temporary file processing
 - [Static IP Address](#)
5. Up-to-date modern browsers for viewing Tricefy on a computer or mobile device
 - a. Supported browsers
 - Chrome
 - Firefox
 - Edge
 - Safari
 - b. Using an out-of-date browser will redirect to the mobile version of Tricefy for security purposes

Note: Internet Explorer is no longer supported as of April 2022

 Learn more: tricefy.help/help/ie
