




## Voluson Series with Tricefy Inside

Last Modified on 2024-05-28 18:06

GE Voluson ultrasound systems have Tricefy pre-installed. Additional software is not necessary in order to use Tricefy. Activate your account using the following steps:

- 1) Press the **Utilities** button on the front panel (this is a physical button), and select **Setup** on the touch screen
- 2) Select **Connectivity**
- 3) Select the **Tricefy** tab and activate your account.

*Note: The Tricefy tab is only available on new machines that have Tricefy Inside.*



**A** - Enter your email address to authorize the account. This should be an IT or clinic administrator.

**B** - Select **Activate Account**. This process may take up to 20 seconds.

- A **green** message means the activation was successful
- An **orange** message the activation failed
  - If activation fails, confirm there are not any firewalls or routers blocking Port 443
  - If activation continues to fail, select the **Contact Tricefy Help Center** button at the bottom of this screen and we will contact you shortly to assist in getting your account activated


**C** - Select **Test Connection** to confirm your ultrasound system can connect to Tricefy

- A **green** message means the test was successful and you can continue with the steps below
- An **orange** message means the ultrasound system failed to connect to Tricefy
  - If the test fails, confirm there are not any firewalls or routers blocking Port 443
  - If the test continues to fail, select the **Contact Tricefy Help Center** button at the bottom of this screen and we will contact you shortly to troubleshoot the issue and get your system connected.

## Sending Images to Tricefy




To program your ultrasound system to send exam data to Tricefy, select the **Button Configuration** tab:



You can program your system to either automatically send exam data to Tricefy once the exam has ended or by selecting a P button. Both options are described below.


## Sending Data at the End of Exam

To send data when the exam has ended, select the **End Exam** option and make sure **Move Clipboard Content to Interval Archive** and **Store Clipboard Content to Tricefy** are selected in the Clipboard Content box.



## Sending Data with a P Button

To send data by pressing a P button, select the P button you would like to use (P1, P2, P3, P4, P5 or P6) and make sure **Save to Clipboard** is selected in the General box. Select **Store Image/Cine** and **Share Image/Cine** in the Tricefy box.




To learn more about sending studies to patients and doctors, refer to [Step 3 of our Getting Started Guide](#).

## Sending Worksheets and Reports

Refer to [tricefy.help/help/pdfsendvol](http://tricefy.help/help/pdfsendvol) to learn how to send worksheets and reports to Tricefy in one easy step.

## Image/Cine Selection

To enable sending both still images and cines with one button, we recommend the following settings under **Detailed Setup**:



- 2D: Automatic
- D/M: Single
- 3D: Automatic
- 4D: Automatic

If you have additional ultrasound systems to setup that don't already have Tricefy installed, refer to the following help pages:

<http://www.tricefy.help/help/uplink>

<http://www.tricefy.help/help/configuration>

## Compression Settings:

Confirm your compression settings are correct:

[Compression Settings for Voluson](#)

